Electronic pressure transducer type DT2 and DT2V

Two-wire system 4 ... 20 mA signal output or Three-wire system 0 ... 10 V DC signal output, 1 ... 5 V DC output Connection via plug M12x1; 4-pin conforming IEC 61076-2-101 (draft)

(6

Pressure range $p_{range} = 0 \dots 600$ bar

1. General information, brief description

The electronic pressure transducers type DT2 are thin-film strain gauges utilizing a full bridge. The sensor elements are temperaturecompensated. The amplifying and adjustment of the measuring signal takes place by analogous electronics.

Functional parts are:

Thin-film strain gauge

Analogous amplifier

Connection via plug M12x1; 4-pin conforming IEC 61076-2-101 (draft) Housing made of stainless and plastic Hydraulic connection via tapped journal G 1/4 (BSPP)

Features:

Nom. pressure ranges 100, 250, 400, and 600 bar Accuracy 1% of p_{range} Pressure peak resistant due capillary dampening system (CDS), extreme temperature shocks and vibration proof High long term stability Two-wire system 4 ... 20 mA, burden 700 Ω (at 24 V DC) or Three-wire system 0 ... 10 V DC, $R_B \ge 5 k\Omega$ Three-wire system 1 ... 5 V DC All parts in contact with the fluid made of stainless (DIN 1.4542 and 316L) Housing made of fiber enforced plastic (PBT) Sturdy industrial design Certified EMC Very good price/performance ratio qualifies

1.1 Circuitry DT2-... (Two-wire system, 4 ... 20 mA)



1.2 Circuitry DT(S)2V-... (Three-wire system, 0 ... 10 V DC)



Simplified symbol

Simplified symbol



4..20 mA

The electronic pressure transducers type DT2 can be utilized in almost all areas of industrial hydraulics. Typical applications are test benches, machines, plant construction as well as automation engineering.

The thin film technology enables the long term reliability often demanded for hydraulics.

The EMC enables signal output without interferences even under rough ambient conditions.

The good price/performance ratio qualifies this pressure transducer for middle to high quantity applications, where reliability and economy are a must.



Δ

HYDRAULIK



HAWE HYDRAULIK SE

STREITFELDSTR. 25 • 81673 MÜNCHEN



Simplified symbol





2. Available versions, type coding key, accessories Pressure transducer with current signal with voltage signal Parts No.: bar Parts No .: bar Order coding: DT2-1 100 6217 8099 **DT2V-1** 100 6217 8103 DT2-2 250 6217 8100 **DT2V-2** 250 6217 8104 **DT2-4** 400 6217 8101 DT2V-4 400 6217 8105 DTS2V-4 DT2-6 600 6217 8102 400 6217 8188 DTS2V4-4 400 6217 8193 DT2V-6 600 6217 8106 Nom. pressure range in bar -Mounting accessories Order codina: K 1/4 Short prolongation G 1/4 - G 1/4 A, with fitting seal ring G 1/4 NBR (BSPP) L 1/4 Order coding: Long prolongation G 1/4 - G 1/4 A, with fitting seal ring G 1/4 NBR (BSPP) Order coding: Plug M12x1; 4-pin MSD-T7, parts No. 6217 8048 3. **Technical data** 3.1 General data Electronic pressure transducer Nomenclature G 1/4 A (BSPP) conf. DIN 3852 E, with NBR seal ring Hydraulic connection Materials in contact with DIN 1.4542 or 316L the pressure fluid Fiber enforced plastic (PBT) Housing materials Version DTS2V with stainless housing Electrical connection Via plug M12x1; 4-pin conforming EN 61076-2-101 (not scope of delivery) Installed position Any approx. 70 g Mass (weight)

compatibility (EMC)	conf. EC regulation 89/336 EWG
Vibration resistance acc. to DIN IEC 68-2	20 g
Shock resistance acc. to DIN EN 837	500 g

3.2 Hydraulic parameters

Protection class DIN EN 60529

or IEC 60529 properly installed

Ambient temperature Compensated range

Fluid temperature

Electro magnetic

			DT2V-1 DT2-1	DT2V-2 DT2-2	DTS2V4-4 DT(S)2V-4 DT2-4	DT2V-6 DT2-6
Measuring range	p _{range}	[bar]	0 100	0 250	0 400	0 600
Perm. pressure overbad	p _{max}	[bar]	200	500	800	1200
Burst pressure	p _{burst}	[bar]	800	1200	1700	2400

IP 67

0°C ... 80°C

-40°C ... +125°C

-40°C ... +100°C (storage -40°C ... +120°C)

Interference emission and interference immunity acc. to EN 61326

Note: The device won't be harmed between p_{range} and p_{max}

But it will be damaged in the range between pmax and pburst but remains tight to the outside.

Mounting accessories K 1/4 and L 1/4:

Max. operation pressure	p _{oper}	[bar]	1000
Burst pressure	p _{burst}	[bar]	approx. 2 x p _{oper}

3.3	Electrical data		
	Accuracy: Linearity within the compensated pressure rang Repeatability	ge	$\pm 1\%$ of the span (typical $\pm 0.6\%)$ $\pm 0.2\%$ of the span
	Temperature drift: Mean TC of zero Mean TC of the span 1-year stability		$\pm 0.15\%$ of the span / 10K $\pm 0.15\%$ of the span / 10K $\pm 0.2\%$ of the span (at reference conditions)
3.3.1	Pressure transducer type DT2 (4 20 mA)		
	Supply voltage	U _B	10 36 V DC Reverse polarity
	Max. perm. ripple factor	W	10% (ripple)
	The supply voltage U _B must be minimum 10 V DC higher than the voltage drop of the external measuring device (see permissible burden R _A and sect. 1.1 "Circuitry").		
	Output:		
	Output signal	I _A	4 20 mA, two-wire system
	Perm. burden	R _A	$R_{A}[\Omega] \leq (U_{B}[V] - 10 V) / 0.02 A$
	Response time (1090%)	t _A	≤ 2 ms
3.3.2	Pressure transducer type	DT2V (0 10 V DC)
	Supply voltage	U _B	14 36 V DC Reverse polarity
	Max. perm. ripple factor	W	10% (ripple)
	Output:		
	Output signal	U _A	0 10 V DC, three-wire system, short-circuit proof
	Perm. burden	R₄	≥5 kΩ
	Response time (1090%)	t _A	≤ 2 ms
3.3.3	Pressure transducer type DTS2V4-4 (1 5 V DC)		
	Supply voltage	U _B	10 30 V DC Reverse polarity
	Max. perm. ripple factor	W	10% (ripple)
	Output:		
	Output signal	U _A	1 5 V DC, three-wire system
	Perm. burden	R _A	\geq 5 k Ω
	Response time (1090%)	t _A	≤ 2 ms

3.4 Electro-magnetic compatibility (EMC)

The EMC of the device was checked by an accredited approval institute (Interference emission and interference immunity acc. to EN 61326). This EMC test doesn't relieve the user from the proper execution of a specified EMC test for his complete system, since these test assemblies represent only a typical application (conforming the EC-guideline 89/336/EWG). The following measures increase the EMC:

- The transducer should be grounded
- The device should be installed in a closed metal cabinet (shielding).
- All cables, leading in or out of the device should be kept as short as possible. They should be also shielded and twisted in pairs. (This will reduce the antenna effect and increase the interference immunity).

3.5 Indications for assembly and initial operation

The electronic pressure transducer fulfills protection class IP 67 only when the plug with connection cable is correctly installed. Moisture might intrude when a too thin cable is used or other leaking spots are apparent. This moisture will get to the internal electronics leading to corrosion and malfunction of the device.

Attention: All defects of the device caused by above causes will harm the warranty!

